



Certificate of Compliance

Certificate: 1091908

Master Contract: 181566

Project: 2351636

Date Issued: February 28, 2011

Issued to: Delta C Technologies Incorporated

506B Granville Ave
Enderby, BC V0E 1V0
Canada
Attention: Vince Vaniderstine

The products listed below are eligible to bear the CSA Mark shown



Rob Kohuch

Issued by: Rob Kohuch, P.Eng.

PRODUCTS

CLASS 2258 02 - PROCESS CONTROL EQUIPMENT - For Hazardous Locations

Class I, Division 1, Groups C and D:

Oil/Water Analyzer, Models DC-1000, DC-1500, DC-1510, DC-3500, DC-3510F, DC-3510NPT;

input rated 24 V dc, 150 mA;

alarm contacts rated 24 V dc, 2 Amp max;

output signal 4-20 mA.

operating ambient 70°C max.,

working pressure 15 MPa (2,250 psig) max.

APPLICABLE REQUIREMENTS

CSA C22.2 No. 0-10 - General Requirements, Canadian Electrical Code, Part II.

CSA C22.2 No. 0.4-04 - Bonding of Electrical Equipment

CSA C22.2 No. 30-M1986 - Explosion-Proof Enclosures for Use in Class I Hazardous Locations.

CSA C22.2 No. 142-M1987 - Process Control Equipment.